

External-Cavity Quantum Cascade Laser Spectroscopy for Mid-IR Transmission Measurements of Proteins in Aqueous Solution


Mirta R. Alcaráz,^{†,‡} Andreas Schwaighofer,[†] Christian Kristament,[†] Georg Ramer,[†] Markus Brandstetter,^{†,§} Héctor Goicoechea,[‡] and Bernhard Lendl^{*,†}

[†]Institute of Chemical Technologies and Analytics, Vienna University of Technology, Getreidemarkt 9/164-UPA, 1060 Vienna, Austria

[‡]Laboratorio de Desarrollo Analítico y Quimiometría (LADAQ), Cátedra de Química Analítica I, Facultad de Bioquímica y Ciencias Biológicas, Universidad Nacional del Litoral-CONICET, Ciudad Universitaria, 3000 Santa Fe, Argentina

Supporting Information

ABSTRACT: In this work, we report mid-IR transmission measurements of the protein amide I band in aqueous solution at large optical paths. A tunable external-cavity quantum cascade laser (EC-QCL) operated in pulsed mode at room temperature allowed one to apply a path length of up to 38 μm , which is four times larger than that applicable with conventional FT-IR spectrometers. To minimize temperature-induced variations caused by background absorption of the ν_2 -vibration of water (HOH-bending) overlapping with the amide I region, a highly stable temperature control unit with relative temperature stability within 0.005 $^\circ\text{C}$ was developed. An advanced data processing protocol was established to overcome fluctuations in the fine structure of the emission curve that are inherent to the employed EC-QCL due to its mechanical instabilities. To allow for wavenumber accuracy, a spectral calibration method has been elaborated to reference the acquired IR spectra to the absolute positions of the water vapor absorption bands. Employing this setup, characteristic spectral features of five well-studied proteins exhibiting different secondary structures could be measured at concentrations as low as 2.5 mg mL^{-1} . This concentration range could previously only be accessed by IR measurements in D_2O . Mathematical evaluation of the spectral overlap and comparison of second derivative spectra confirm excellent agreement of the QCL transmission measurements with protein spectra acquired by FT-IR spectroscopy. This proves the potential of the applied setup to monitor secondary structure changes of proteins in aqueous solution at extended optical path lengths, which allow experiments in flow through configuration.



Infrared spectroscopy is a versatile technique for the quantitative and qualitative analysis of a wide range of samples. After the development of the fast Fourier transform algorithm in the 1960s,¹ it was feasible to employ the Michelson-interferometer for IR spectroscopy. Since then, Fourier transform infrared (FT-IR) spectrometers have been established as standard instrumentation. Thermal emitters such as globars are commonly used as light sources that emit a broadband but low-power radiation throughout the mid-IR region. While these characteristics suffice for a variety of applications, the low emission intensity of the light source constitutes a limit for the measurement of samples with a highly absorbing matrix, e.g., water. In these cases, the abundant solvent absorbs a large part of the irradiated light, thus attenuating the light intensity which is available for interaction with the actual analyte. As a consequence, the maximum feasible path length for transmission measurements in aqueous solution is limited in the carbohydrate region (950–1200 cm^{-1}) to approximately 80 μm and in the amide I region of proteins (1600–1700 cm^{-1}) to approximately 8 μm .^{2,3} These

low path lengths may considerably impair the robustness of the method due to clogging of the cell.

Two decades ago, a new type of mid-IR light source, quantum cascade lasers (QCLs), was introduced.⁴ QCLs provide spectral power densities that are several orders of magnitude higher than globars and even surpass values of synchrotrons.⁵ QCLs comprise a series of layers of different semiconductor materials exhibiting varying band gaps with a defined thickness in the nanometer range. Laser emission is achieved by inter sub-band transitions of electrons within the semiconductor conduction band. Only one type of charge carrier (electrons) is used contrary to conventional semiconductor lasers, where electrons and holes from the conduction and valence band combine in a radiative interband transition. In conventional lasers, the emission wavelength is determined by the band gap of the utilized materials, thus restricting their application to the UV–vis and near-IR range.

Received: May 7, 2015

Accepted: June 10, 2015

In QCLs, the emission wavelength is decoupled from the band gap of the semiconductor material and lasers emitting at a desired wavelength region can be engineered by tuning the thicknesses of the semiconductor layers.^{6–8} Initially, QCLs were predominantly used for gas phase analysis.^{9–12} Distributed feedback (DFB) QCLs feature a grating on top of the active semiconductor material for selection of the desired wavenumber. Limited tunability across a few wavenumbers is adequate for gas phase analysis, where spectra are dominated by narrow, well resolved bands. This is not the case for the investigation of liquid samples, where absorption bands are much broader and overlapping of spectral components may occur, too. Hence, for the analysis of liquids, a broader spectral range is essential. Recently, external-cavity QCLs (EC-QCLs), that are operated at room temperature and combine high spectral power densities with a large spectral tuning range, became commercially available. Here, an external cavity facilitates wavenumber selection by an external diffraction grating. Changing the angle of the diffraction grating relative to the QCL chip allows for one to tune ranges of up to several hundred wavenumbers. This type of QCL was applied for analysis of complex mixtures of analytes in aqueous solution in online process monitoring¹³ and for medical applications.^{2,14–16} High emission intensities made it possible to employ a larger path length for liquid transmission measurements and enhance the ruggedness of the resulting analysis system compared to FT-IR spectroscopy.

Infrared spectroscopy is a well-established experimental technique for the secondary structure analysis of proteins.¹⁷ The most prominent absorption bands of proteins in the IR spectrum are caused by vibrations of the peptide group and are referred to as amide bands. Among these characteristic bands in the mid-IR region, the amide I band (1600–1700 cm^{-1}), which is mainly composed of C=O stretching vibrations, is most commonly used for analysis of the secondary structure.¹⁸ The differing pattern of hydrogen bonding, dipole–dipole interactions, and geometric orientations in the α -helices, β -sheets, turns, and random coil structures induce different frequencies of the C=O vibrations that can be correlated with the respective secondary structural folding.¹⁹ Quantitative estimation of protein secondary structure has often been performed by curve fitting of the amide I band. This approach, however, often lacks accuracy, and its applicability is limited as it requires a series of subjective decisions^{18,20} and has a tendency to over- or underestimate secondary structure components.^{21,22} Transmission measurements are most frequently used for recording solution spectra of proteins, while the attenuated total reflectance (ATR) method is employed for adsorption studies or investigation of thin films.

The most important difficulty of IR investigations of proteins in aqueous solution is the strong absorbance of the HOH-bending band near 1645 cm^{-1} that overlaps with the amide I band. Consequently, suitable path lengths are restricted to approximately 8 μm to prevent total IR absorption in the region of the HOH-bending band, as outlined above. Thus, high protein concentrations (>10 mg mL^{-1}) are required as the intensities of the IR bands and the signal-to-noise ratio at a given concentration is limited.¹⁸ Cell and sample handling under these conditions is rather laborious as thin spacers are extremely susceptible to electrostatic charging making the assembly of a tight flow cell difficult. Further, highly concentrated protein solutions are viscous leading to the formation of air bubbles when filling the cell.³ In this context, a

further experimental challenge for obtaining appropriate protein spectra is to ensure experimental conditions that are identical for the background and sample measurement. Since changing temperatures cause variations in the HOH-bending band, temperatures of the aqueous buffer and sample solution should match within at least 0.1 $^{\circ}\text{C}$ in order to avoid artifacts caused by temperature differences.

Alternatively, D_2O is widely used as solvent for protein measurements. The DOD-bending band is located at $\sim 1200 \text{ cm}^{-1}$, thus creating a region of relatively low solvent absorbance between 1500 and 1800 cm^{-1} . This allows for larger path length ($\sim 50 \mu\text{m}$) and lower protein concentrations ($\sim 1 \text{ mg mL}^{-1}$). Solvent exchange ($\text{H}_2\text{O} \rightarrow \text{D}_2\text{O}$) leads to a shift of the amide I band to lower wavenumbers due to the replacement of hydrogen atoms in the protein backbone by heavier deuterium atoms.^{18,23} Despite the experimental advantages offered by D_2O , it is still preferable to use H_2O when studying the protein structure, as it provides the native environment.^{3,24,25} The effect on H–D exchange on protein structure is not fully resolved, but it has been reported that exposure of proteins to D_2O may alter the strength and length of hydrogen bonds and leads to changes in structure and protein dynamics²⁶ as well as protein denaturation.^{27,28}

In this work, we introduce a rugged setup based on an EC-QCL for mid-IR transmission measurements of aqueous protein solutions in the amide I region. The high emission intensity of the tunable QCLs enabled one to increase the path length up to 38 μm . A custom-built temperature control unit was designed that provides a relative temperature accuracy of the measurement cell in the order of 0.005 $^{\circ}\text{C}$ to minimize temperature-induced variations of the HOH-bending band in the spectrum. The focus of this paper is to demonstrate the feasibility of this setup to elucidate the spectral features of different protein secondary structures. For this purpose, spectra of five proteins with well-known secondary structures were recorded with a QCL-based transmission setup and contrasted with FT-IR spectra. Evaluation by second-derivative spectra confirmed excellent comparability of the acquired spectra.

EXPERIMENTAL SECTION

Reagents and Samples. Sodium phosphate monobasic dihydrate p.a. ($\text{NaH}_2\text{PO}_4 \cdot 2\text{H}_2\text{O}$) was purchased from Fluka (Buchs, Switzerland); sodium phosphate dibasic dihydrate ($\text{Na}_2\text{HPO}_4 \cdot 2\text{H}_2\text{O}$) BioUltra, for molecular biology, and sodium hydroxide solution 50% in water were obtained from Sigma-Aldrich (Steinheim, Germany). β -Lactoglobulin from bovine milk ($\geq 85\%$), hemoglobin from bovine blood, bovine serum albumin ($\geq 98.0\%$), lysozyme from chicken egg white ($>90\%$), and α -chymotrypsin from bovine pancreas ($\geq 85\%$) were obtained from Sigma-Aldrich (Steinheim, Germany) and used as purchased. An appropriate amount of lyophilized protein powder was dissolved in 16.0 mmol L^{-1} phosphate buffer pH 7.0. Ultrapure water (18 $\text{M}\Omega$) was used for the preparation of all solutions, obtained with a Milli-Q water purification system from Millipore (Bedford, USA).

FT-IR Measurements. FT-IR absorption measurements were performed using a Bruker 80v FT-IR spectrometer (Ettlingen, Germany) equipped with a liquid nitrogen cooled HgCdTe (Mercury cadmium Telluride) detector ($D^* = 4 \times 10^{10} \text{ cm Hz}^{0.5} \text{ W}^{-1}$ at 9.2 μm). The samples were placed between two CaF_2 windows separated by an 8 μm -thick spacer. During measurements, the spectrometer was flushed with dry air for at least 10 min prior to spectrum acquisition.

Spectra were acquired with a spectral resolution of 2 cm^{-1} in double-sided acquisition mode. A total of 32 scans were averaged per spectrum (acquisition time: 6.7 s), which was calculated using a Blackman-Harris 3-term apodization function and a zero filling factor of 2. All spectra were acquired at $25\text{ }^{\circ}\text{C}$. Spectra were analyzed using the software package OPUS 7.2 (Bruker, Ettlingen, Germany). If necessary, absorption bands of water vapor in the atmosphere were subtracted.

Experimental Setup for QCL Measurements. The experimental setup is shown in Figure 1. For the measurements,




Figure 1. Representation of the experimental QCL-based setup for mid-IR transmission measurements.

a thermoelectrically cooled external-cavity quantum cascade laser (Daylight Solutions Inc., San Diego, USA) was used operating at a repetition rate of 100 kHz and a pulse width of 500 ns. The manufacturer states laser power variations of $<2\%$ RMS and an operating temperature range of $15\text{--}35\text{ }^{\circ}\text{C}$. The laser head temperature was set to $18\text{ }^{\circ}\text{C}$ for all measurements. All spectra were recorded in the spectral tuning range located between 1729.30 and 1565.06 cm^{-1} , covering the amide I region of proteins. The relation between the spectral tuning time and the corresponding emission wavenumber was determined through characterization of laser tuning using the step-scan capability of the employed FT-IR spectrometer.¹⁵ The MIR light was focused on the detector element by a gold plated off-axis parabolic mirror with a focal length of 43 mm. A thermoelectrically cooled MCT detector operating at $-60\text{ }^{\circ}\text{C}$ (Infrared Associates Inc., USA; MCT-7-TE3) with a $1 \times 1\text{ mm}$ element size and a detectivity of $D^* = 4 \times 10^9\text{ cm Hz}^{0.5}\text{ W}^{-1}$ at $9.2\text{ }\mu\text{m}$ was used as IR detector.

The measured signal was processed by a two channel boxcar integrator digitized by a NI DAQ 9239 24-bit ADC (National Instruments Corp., Austin, USA). Each single beam spectrum consisting of 24 000 data points was recorded during the tuning time of 1.5 s. A total of 50 scans were recorded for background and sample single beam spectra. An AD7760 24-bit ADC (Analog Devices, Norwood, USA) was employed to investigate the relative wavenumber shifting.

The spectral resolution of the EC-QCL setup was determined by evaluation of the bandwidth of water vapor spectra and comparison with FT-IR spectra acquired at different resolutions. After applying scan-to-scan alignment and background-sample alignment (see next section), the resolution of EC-QCL spectra was determined to be 0.2 and 1.2 cm^{-1} for nonfiltered and filtered spectra, respectively.

All measurements were carried out using a custom-made, temperature-controlled flow cell equipped with two MIR transparent CaF_2 windows and a $38\text{ }\mu\text{m}$ -thick spacer, at approximately $24.6\text{ }^{\circ}\text{C}$ with relative temperature stability within $0.005\text{ }^{\circ}\text{C}$. The temperature cell consists of eight thermoelectric cooling (TEC) elements, stabilized by liquid water cooling and able to be operated from 0 to $80\text{ }^{\circ}\text{C}$. A sensor is located near the CaF_2 windows to provide constant monitoring of the temperature. The cell is equipped with a meander (1565 mm) to provide temperature stabilization for flow through measurements. To reduce the influence of water vapor, the setup was placed in a housing of polyethylene foil and constantly flushed with dry air. Automatic sampling was done by a flow injection system, consisting of a 14-way valve (Valco Instruments Co. Inc., Schenkon, Switzerland), a 2.5 mL syringe pump (Cavro XP3000 Tecan Systems Inc., Männedorf, Switzerland), and PTFE tubings in several dimensions. The whole setup was controlled by a LabView-based GUI (National Instruments Corp., Austin, USA) with server–client program structure.²⁹ The dimensions of the setup including electronic parts are $58\text{ cm} \times 47\text{ cm} \times 77\text{ cm}$ ($L \times W \times H$).

Processing of QCL Data. The performed preprocessing steps are schematically depicted in Figure 2. First, the number




Figure 2. Sequence of processing steps for EC-QCL IR transmission spectra.

of data points in the single beam was reduced by a factor of 4 using cubic spline interpolation followed by Savitzky-Golay smoothing to reduce instrumental noise. To correct the spectral mismatch of successive scans (see discussion for details), Correlation Optimized Warping (COW) was applied to align consecutive scans of one measurement prior to averaging as well as to align the background with the sample single beam spectrum. Scan-to-scan alignment was performed to 50 single beam spectra, taking the middle spectrum as reference. After alignment, the spectra were averaged. Background (I_0) – sample (I) alignment was applied taking the background-spectrum as reference. COW is a widely used method to correct spectral (NMR, IR, or Raman) or chromatographic shifts. Two parameters have to be set, the segment size r and a slack size l . Considering two spectra, one is chosen as the reference. The




Figure 3. IR absorbance spectra of 10 mg mL⁻¹ (green solid line), 5 mg mL⁻¹ (green dashed line), and 2.5 mg mL⁻¹ (green dotted line) protein solutions acquired by the EC-QCL setup (A–E) and 20 mg mL⁻¹ (brown solid line) protein solutions acquired by FT-IR spectroscopy (F–J). Gray dashed lines highlight the high congruence of the spectral features between the IR spectra acquired by EC-QCL and FT-IR spectroscopy. Gray double-headed arrows indicate the absorbance of 10 mAU.

second spectrum is subdivided into sections of length r which are iteratively stretched and compressed for maximal l data points via interpolation so as to maximize the correlation between the spectra. This kind of alignment distorts neither the amplitude nor the wavenumber position of the final absorbance spectrum.³⁰ In this work, r was kept constant at 20 data points (corresponding to 0.45 cm⁻¹) and selected by considering the length of mode-hops structure, and l was selected and optimized within the range of 2–5 data points for each sample.

After alignment of the sample and background spectrum, a low-pass Fourier filter based on 4-term Blackman-Harris apodization (fast Fourier transformation, FFT) was applied to single beam spectra I and I_0 . The final absorption spectrum was obtained using

$$A = -\log\left(\frac{I}{I_0}\right) \quad (1)$$

In order to quantitatively evaluate the comparability of the protein IR spectra acquired by EC-QCL and FT-IR spectroscopy by the degree of spectral overlap (s_{12}) between FT-IR (s_1) and EC-QCL (s_2), the following expression was employed:³¹

$$s_{12} = \frac{\|s_1^T s_2\|}{\|s_1\| \|s_2\|} \quad (2)$$

This evaluation was performed for absorbance spectra as well as second derivative spectra. The value of s_{12} ranges from 0 to 1, corresponding to no overlapping and complete overlapping, respectively.

Data processing and analysis was performed with in-house code developed in MatLab R2014b (MathWorks, Inc., Natick, MA, 2014). The COW routine is available online at http://models.kvl.dk/DTW_COW.

RESULTS AND DISCUSSION

Mid-IR Spectra Recorded with the EC-QCL Setup. Mid-IR transmission spectra were recorded of aqueous solutions of five proteins at different concentrations to assess the capabilities of the EC-QCL setup for monitoring protein secondary structure (see Figure 3A–F). The broad tuning range of the applied EC-QCL allows one to record spectra across the amide I range, thus enabling quantitative as well as qualitative evaluation of the protein spectra. The investigated proteins exhibit characteristic spectral features of different secondary structures. The spectra of bovine serum albumin (BSA) and hemoglobin (Hemo), two well-studied proteins consisting primarily of α -helical structures,^{32,33} show the α -helix characteristic band at 1656 cm⁻¹.¹⁷ α -Chymotrypsin (ACT) and β -lactoglobulin (BLG) are mainly composed of β -sheet secondary structure with a characteristic band maximum at approximately 1635 cm⁻¹ and a sideband at 1680 cm⁻¹.^{32,34} Lysozyme (Lys) features both, α -helices and β -sheets, resulting in a band maximum at 1656 cm⁻¹ with shoulders at \sim 1640 and \sim 1675 cm⁻¹.^{32,35} Spectral features of the diverse secondary structures could be identified at a concentration as low as 2.5 mg mL⁻¹. This could be achieved by employing a high intensity EC-QCL enabling the increased optical length of 38 μ m conjoined with a sophisticated data processing routine.

Band areas were evaluated for proteins solutions with five concentrations of every sample. The calibration curves in Figure S-1, Supporting Information, show straight lines and demonstrate linearity down to a concentration of 1 mg mL⁻¹. The validity of the Lambert–Beer law is significant for prospective applications of this setup for quantitative analysis.

Processing of Raw Data. In EC-QCLs, spectral tuning across the wavenumber range is achieved by a quasi-continuous rotation of the EC grating. The resulting emission curve features a fine structure originating from the mode-hops of the laser chip and external cavity.³⁶ Mechanical imperfections as well as triggering issues lead to a spectral mismatch in the fine

structure of consecutive scans (as shown in Figure S-2A, Supporting Information). The fluctuations appear small in the single beam spectra but cause considerable noise in the corresponding absorbance spectrum.³⁷ Since these fluctuations are not random, averaging of a high number of scans is not an option to decrease its disturbance. It had been attempted to reduce the effect of the fluctuations on the absorbance spectrum by shifting the recorded scans along the wavenumber axis to achieve maximum alignment.³⁷ However, since the shift is not constant during one scan, even after applying this procedure, extensive use of Fourier filtering was necessary to obtain optimal spectrum quality. Fourier filtering tends to introduce distortions in the line shape, thus affecting the information contained in the amide I band.³⁸ To keep usage of Fourier filtering to a minimum, an advanced spectral alignment routine was elaborated (as depicted in Figure 2). The key element of this sequence is the COW algorithm applied here to align scans prior to averaging as well as to align the background with the sample spectrum. This method utilizes the EC-QCL inherent mode-hop fine structure for alignment. It not only shifts the spectra in the wavenumber axis but also performs a rubberband type alignment. This part turned out to be crucial since the mismatch of the fine structure is not constant throughout the spectrum. A comparison between the shift- and COW-based alignment is shown in Figure S-2B, Supporting Information. Noise reduction by shift alignment was insufficient for protein spectra. With this method, spectral features of the respective secondary structures could be recognized for protein concentrations of 10 mg mL⁻¹. By introducing the COW procedure, the identification of characteristic spectral features of diverse secondary structures could be accomplished for concentrations as low as 2.5 mg mL⁻¹. This improvement is mainly credited to the significantly improved reduction of noise. In consequence, merely minor Fourier filtering is required, with a cutoff frequency of 50 Hz.

Wavenumber Calibration of the EC-QCL Setup. The accuracy of the wavenumber scale in the IR spectra has been the topic of discussion for spectroscopic instruments³⁹ and in particular for QCL setups.³⁷ Deviations in the wavenumber scale may occur from inaccuracies in the light source and data acquisition as well as data processing. In EC-QCLs, one sweep of the grating in a continuous motion results in the wavenumber scan. The calibration of the wavenumber axis with tuning time for the employed EC-QCL is performed through characterization with a time-resolved (step-scan) FT-IR measurement.¹⁵ Regarding data acquisition, it was found that the type of ADC has a considerable effect on the resulting IR spectrum, due to different time lags owing to triggering inaccuracies and internal data processing routines. As outlined above, the temporal axis directly corresponds to the wavenumber axis; thus, time lags introduced in data acquisition translate into wavenumber deviations.

In order to correct the deviations in the wavenumber scale, the EC-QCL setup has been calibrated to the absorption lines of water vapor. Previously, a method for frequency calibration of EC-QCLs has been proposed that utilizes the inherent fine-structure superimposed to the laser emission curve.³⁷ Here, high resolution FT-IR spectra of the EC-QC laser emission are recorded at fixed grating positions; thus, the FT-IR spectrometer serves as an absolute wavelength reference. A notable advantage of the calibration method employing the absorption bands of the ubiquitous water vapor as a reference is the independence from external equipment. Frequency

calibration to gas-phase absorption bands has been widely applied for different kinds of spectrometers, and reference data for various gases is available from the HITRAN database.^{39,40} Using the EC-QCL setup, IR spectra of water vapor have been recorded and the absorption lines were compared to reference data. Figure S-3A, Supporting Information, shows the shift of water vapor bands acquired with two available ADCs relative to the position of vapor bands taken from the reference database. The results indicate that the ADCs introduce a characteristic and nonconstant shift to the wavenumber scale. For the AD7760, the deviation is about -3 cm^{-1} across the observed wavenumber region, with a small slope ($k = -0.0083$), while for the NI ADC, the aberration is generally smaller (-1 to $+3\text{ cm}^{-1}$) but with a larger change across the observed wavenumber region ($k = 0.0408$). This behavior did not change for repeated measurements. Due to a lower noise level ($\text{RMS}_{\text{AD7760}} = 0.54\text{ mAU}$, $\text{RMS}_{\text{NI}} = 0.37\text{ mAU}$), the NI ADC was used for protein measurements. The relation between the recorded and the reference wavenumber scale obtained by the water vapor spectrum was used to calibrate the protein spectra. Figure S-3B, Supporting Information, shows the efficiency of the wavenumber calibration for a water vapor band. This frequency calibration accounts for all deviations introduced in the experimental setup (EC-QCL, data acquisition). The remaining wavenumber deviation introduced by shifting and warping of the COW processing routine was found to be lower than 0.1 cm^{-1} which is negligible for liquid phase IR measurements.

Enabling Temperature-Stabilized Measurements. The ν_2 -vibration of water (HOH-bending) with a maximum near 1645 cm^{-1} overlaps with the amide I band of proteins.⁴¹ The position of the bending band is rather sensitive to changes of the temperature, as it depends on the strength of the hydrogen bonds. Upon a temperature increase, the hydrogen bonds become weaker, leading to a shift of the bending vibration to lower wavenumbers accompanied by narrowing of the band and an increase of intensity.⁴² This behavior results in a rather complex change of the solvent spectrum.⁴³ The impact is particularly strong due to the high abundance of water as a solvent and its high molar absorptivity in this region.⁴¹ Further, it has been reported that altering the temperatures also affects the effective path length of the transmission cell due to the thermal expansion of the spacer material.⁴⁴ Thermal expansion of a $38\text{ }\mu\text{m}$ Teflon spacer (linear expansion coefficient $\alpha = 0.00012\text{ K}^{-1}$)⁴⁵ at a temperature change of 1 K leads to an increase of the path length of $0.005\text{ }\mu\text{m}$, resulting in a change of absorbance of 0.57 mAU for the HOH bending band ($c = 55.56\text{ mol L}^{-1}$, $\epsilon = 22.3086\text{ L mol}^{-1}\text{ cm}^{-1}$ at 1645 cm^{-1}). The combination of these effects leads to a change of shape and intensity in the resulting absorbance spectrum, in case the background and sample single beam spectra are recorded at different temperatures. Since, for the qualitative assessment of spectral features attributed to protein secondary structure, the band shape and relative intensities are important, measurements performed at constant temperatures are essential. To achieve this, a highly constant temperature cell has been designed. The temperature stability of the employed cell lies within $0.005\text{ }^\circ\text{C}$ across a time period of 30 min.

Comparison of Mid-IR Spectra Recorded with the EC-QCL Setup and a FT-IR Spectrometer. Figure S-4, Supporting Information, shows the 100% lines of water for the EC-QCL setup and the FT-IR instrument at acquisition parameters used for protein measurements. 100% lines are employed for evaluation of the noise level of a spectrometer




Figure 4. Inverted second derivative spectra of sample proteins acquired by the EC-QCL setup (A–E) and FT-IR spectroscopy (F–J). Gray dashed lines highlight the high overlap of the spectral features between the IR spectra acquired by EC-QCL and FT-IR spectroscopy.

setup and are obtained by calculating the absorbance spectrum of two subsequent single beam spectra of the same sample under identical conditions. Under ideal conditions, the result would be a flat line at 100% transmittance, corresponding to zero absorbance.¹⁵ The RMS (root-mean-square) of the 100% line at path lengths employed for protein measurements is 0.05 mAU for the FT-IR setup and 0.37 mAU for the EC-QCL setup. When considering the approximately four times larger absorbance values gained by utilizing a cell with 38 μm path length for the EC-QCL setup compared to 8 μm in the FT-IR spectrometer, the resultant signal-to-noise ratios (SNR) are in the same order of magnitude (e.g., $\text{SNR}_{\text{FT-IR}} \sim 820$ and $\text{SNR}_{\text{EC-QCL}} \sim 550$ for 20 mg mL^{-1} protein solutions). Moreover, for assessing the noise level of a measurement, all contributors need to be considered. Regarding the light source, the noise level introduced by low intensity thermal light sources in FT-IR spectrometers can usually be neglected. High intensity laser sources, however, contribute to the overall measurement noise, particularly when operated in pulsed mode. Referring to the detector, the detectivity of the LN_2 cooled MCT-detector in the FT-IR spectrometer is ten times higher than the Peltier-cooled MCT-detector applied in the QCL setup. An important point to consider for a transmission setup is the employed path length. Generally, when sufficient light is passing through the sample and reaching the detector (i.e., low path length), the noise level will essentially remain constant. However, when increasing the path length up to a range where most of the light intensity is absorbed by the sample, the measurement noise will begin to increase disproportionately, as demonstrated by the 100% line of water recorded by the FT-IR spectrometer utilizing the cell with a path length of 38 μm (RMS = 99.94 mAU), as demonstrated in Figure S-4, Supporting Information. Accordingly, the increase of the path length accompanied by the gain in ruggedness of the analysis system is exclusive to the EC-QCL setup while featuring a comparable SNR to FT-IR spectroscopy.

FT-IR spectra of the protein solutions were recorded as a reference to demonstrate the equality of the protein IR spectra

acquired by EC-QCL with IR spectra recorded by well-established FT-IR spectroscopy. Figure 3F–J shows the IR spectra of 20 mg mL^{-1} protein solutions recorded with a path length of 8 μm . Figure 3 also shows that the shape and position of the absorption bands recorded by the EC-QCL setup correspond well with the conventional method. In order to quantify the congruence between the (normalized) spectra, the degree of spectral overlap was computed. Using eq 2, the s_{12} values obtained for BSA, Hemo, ACT, BLG, and Lys were 0.99569, 0.99668, 0.99959, 0.99853, and 0.99783, respectively. These figures allow us to conclude that the spectra acquired with the EC-QCL setup contain equivalent information as the FT-IR spectra. For further evaluation, second-derivative spectra are shown in Figure 4. This band narrowing technique is commonly employed to resolve overlapping bands in complex spectral regions.^{3,18} Second-derivative spectra of protein samples acquired by EC-QCL (Figure 3A–E) show similar spectral features as the ones recorded by FT-IR spectroscopy (Figure 3F–J). Also, here, the mathematical evaluation method to quantify the spectral overlap was applied. s_{12} values of 0.99006, 0.97120, 0.98534, 0.96762, and 0.98588 for BSA, Hemo, ACT, BLG, and Lys, respectively, indicate high spectral overlap, considering the enhanced sensitivity of the second-derivative spectra.

CONCLUSIONS AND OUTLOOK

A robust setup employing a broadly tunable EC-QCL was introduced for IR transmission measurements of protein solutions. Characteristic spectral features of five proteins with different secondary structures have been successfully measured at concentrations as low as 2.5 mg mL^{-1} . Mathematical evaluation of the spectral overlap and comparison of second derivative spectra confirm an excellent match of the QCL transmission measurements with protein spectra acquired by FT-IR spectroscopy. The validity of the Lambert–Beer law has been demonstrated for protein concentrations between 1 and 20 mg mL^{-1} . To allow for accuracy in the wavenumber scale employing the EC-QCL setup, a calibration method has been

elaborated to reference the acquired IR spectra to the absolute positions of the water vapor absorption bands. An advanced data processing routine has been devised to diminish intensity variations inherent to the emission spectra EC-QCLs. In the future, arrays of DFB-QCLs that provide broadband wavelength tunability without featuring moving parts may represent an important step to overcome the experimental difficulties encountered with EC-QCLs. The high optical power provided by the QCL allowed one to use significantly higher optical paths (38 μm) than applicable with conventional FT-IR spectrometers. This increase of accessible optical path length opens a wide range of new experimental possibilities in a spectral region that has been limited due to the high absorbance of water. The larger flow cross section results in a greatly reduced pressure drop in the cell and allows experiments in flow through configuration. That signifies a major step forward for IR investigations of proteins, which mostly have been conducted in D_2O , accompanied by more laborious sample preparation and ambiguities regarding the non-native environment. In the future, the introduced setup will be employed to monitor changes of protein secondary structure in aqueous solution induced by heat and chemical denaturation.

■ ASSOCIATED CONTENT

● Supporting Information

Figure S-1: Calibration curves for proteins. Figure S-2: Illustration of processing steps for single beam spectra. Figure S-3: Shift of the water vapor absorption bands recorded with the EC-QCL setup compared to the reference positions taken from a database. Figure S-4: 100% lines for EC-QCL setup and FT-IR instrument. The Supporting Information is available free of charge on the ACS Publications website at DOI: 10.1021/acs.analchem.5b01738.

■ AUTHOR INFORMATION

Corresponding Author

*E-mail: bernhard.lendl@tuwien.ac.at.

Present Address

§M.B.: Research Center for Non Destructive Testing GmbH, Altenbergerstraße 69, 4040 Linz, Austria.

Notes

The authors declare no competing financial interest.

■ ACKNOWLEDGMENTS

The authors thank Dieter Baurecht for fruitful discussions. Financial support was provided by the Austrian research funding association (FFG) under the scope of the COMET programme within the research project “Industrial Methods for Process Analytical Chemistry - From Measurement Technologies to Information Systems (imPACts)” (contract #843546). M.R.A. gratefully acknowledges the financial support provided by CONICET.

■ REFERENCES

- (1) Cooley, J. W.; Tukey, J. W. *Math. Comput.* **1965**, *19*, 297–301.
- (2) Brandstetter, M.; Volgger, L.; Genner, A.; Jungbauer, C.; Lendl, B. *Appl. Phys. B: Lasers Opt.* **2013**, *110*, 233–239.
- (3) Yang, H.; Yang, S.; Kong, J.; Dong, A.; Yu, S. *Nat. Protoc.* **2015**, *10*, 382–396.
- (4) Faist, J.; Capasso, F.; Sivco, D. L.; Sirtori, C.; Hutchinson, A. L.; Cho, A. Y. *Science* **1994**, *264*, 553–556.
- (5) Weida, M. J.; Yee, B. In *Imaging, Manipulation, and Analysis of Biomolecules, Cells, and Tissues IX*; Farkas, D. L., Nicolau, D. V., Leif, R. C., Eds.; Proc. SPIE: San Francisco, 2011; pp 1–7.
- (6) Yao, Y.; Hoffman, A. J.; Gmachl, C. F. *Nat. Photonics* **2012**, *6*, 432–439.
- (7) Capasso, F. *Opt. Eng.* **2010**, *49*, 111102 DOI: 10.1117/1.3505844.
- (8) Curl, R. F.; Capasso, F.; Gmachl, C.; Kosterev, A. A.; McManus, B.; Lewicki, R.; Pusharsky, M.; Wysocki, G.; Tittel, F. K. *Chem. Phys. Lett.* **2010**, *487*, 1–18.
- (9) Weidmann, D.; Tittel, F. K.; Aellen, T.; Beck, M.; Hofstetter, D.; Faist, J.; Blaser, S. *Appl. Phys. B: Lasers Opt.* **2004**, *79*, 907–913.
- (10) Shi, Q.; Nelson, D. D.; McManus, J. B.; Zahniser, M. S.; Parrish, M. E.; Baren, R. E.; Shafer, K. H.; Harward, C. N. *Anal. Chem.* **2003**, *75*, 5180–5190.
- (11) Nagele, M.; Hofstetter, D.; Faist, J.; Sigrüst, M. W. *Anal. Sci.* **2001**, *17*, S497–S499.
- (12) Wysocki, G.; McCurdy, M.; So, S.; Weidmann, D.; Roller, C.; Curl, R. F.; Tittel, F. K. *Appl. Opt.* **2004**, *43*, 6040–6046.
- (13) Siegmann-Hegerfeld, T.; Genner, A.; Brandstetter, M.; Miltner, M.; Lendl, B.; Harasek, M. *Chem. Eng. Trans.* **2013**, *35*, 979–984.
- (14) Brandstetter, M.; Genner, A.; Anic, K.; Lendl, B. *Analyst* **2010**, *135*, 3260–3265.
- (15) Brandstetter, M.; Lendl, B. *Sens. Actuators, B: Chem.* **2012**, *170*, 189–195.
- (16) Brandstetter, M.; Sumalowitsch, T.; Genner, A.; Posch, A. E.; Herwig, C.; Drolz, A.; Fuhrmann, V.; Perkmann, T.; Lendl, B. *Analyst* **2013**, *138*, 4022–4028.
- (17) Barth, A. *Biochim. Biophys. Acta* **2007**, *1767*, 1073–1101.
- (18) Fabian, H.; Mäntele, W. In *Handbook of Vibrational Spectroscopy*; Chalmers, J. M., Griffiths, P. R., Eds.; John Wiley & Sons, Ltd: Chichester, 2002; pp 3399–3425.
- (19) Bal Ram, S. In *Infrared Analysis of Peptides and Proteins*; Bal Ram, S., Ed.; American Chemical Society: Washington, DC, 1999; pp 2–37.
- (20) Goormaghtigh, E.; Ruyschaert, J. M.; Raussens, V. *Biophys. J.* **2006**, *90*, 2946–2957.
- (21) Jackson, M.; Mantsch, H. H. *Crit. Rev. Biochem. Mol. Biol.* **1995**, *30*, 95–120.
- (22) Oberg, K. A.; Ruyschaert, J. M.; Goormaghtigh, E. *Eur. J. Biochem.* **2004**, *271*, 2937–2948.
- (23) Ramer, G.; Balbekova, A.; Schwaighofer, A.; Lendl, B. *Anal. Chem.* **2015**, *87*, 4415–4420.
- (24) Kong, J.; Yu, S. *Acta Biochim. Biophys. Sin.* **2007**, *39*, 549–559.
- (25) Bouhekka, A.; Bürgi, T. *Appl. Surf. Sci.* **2012**, *261*, 369–374.
- (26) Das, D. K.; Mondal, T.; Mandal, U.; Bhattacharyya, K. *ChemPhysChem* **2011**, *12*, 814–822.
- (27) Rial, E.; Muga, A.; Valpuesta, J. M.; Arrondo, J.-L. R.; Goñi, F. M. *Eur. J. Biochem.* **1990**, *188*, 83–89.
- (28) Arrondo, J. L. R.; Goñi, F. M. *Prog. Biophys. Mol. Biol.* **1999**, *72*, 367–405.
- (29) Wagner, C.; Genner, A.; Ramer, G.; Lendl, B. In *Modeling, Programming and Simulations Using LabVIEW*; De Asmundis, R., Ed.; InTech: Rijeka, 2011; pp 1–20.
- (30) Jellema, R. H. In *Comprehensive Chemometrics: Chemical and Biochemical Data Analysis*; Brown, S. D., Tauler, R., Walczak, B., Eds.; Elsevier: Oxford, 2010; pp 85–108.
- (31) Culzoni, M. J.; Goicoechea, H. C.; Ibañez, G. A.; Lozano, V. A.; Marsili, N. R.; Olivieri, A. C.; Pagani, A. P. *Anal. Chim. Acta* **2008**, *614*, 46–57.
- (32) Levitt, M.; Greer, J. J. *Mol. Biol.* **1977**, *114*, 181–239.
- (33) Grdadolink, J.; Marechal, Y. *Biopolymers* **2001**, *62*, 40–53.
- (34) Monaco, H. L.; Zanotti, G.; Spadon, P.; Bolognesi, M.; Sawyer, L.; Eliopoulos, E. E. *J. Mol. Biol.* **1987**, *197*, 695–706.
- (35) Goormaghtigh, E.; Cabiaux, V.; Ruyschaert, J. M. *Eur. J. Biochem.* **1990**, *193*, 409–420.
- (36) Wysocki, G.; Curl, R. F.; Tittel, F. K.; Maulini, R.; Bulliard, J. M.; Faist, J. *Appl. Phys. B: Lasers Opt.* **2005**, *81*, 769–777.

- (37) Brandstetter, M.; Koch, C.; Genner, A.; Lendl, B. In *Quantum Sensing Nanophotonic Devices Xi*; Razeghi, M., Tournié, E., Brown, G. J., Eds.; Proc. SPIE: San Francisco, 2014; pp 1–11.
- (38) Echabe, I.; Encinar, J. A.; Arrondo, J. L. R. *Biospectroscopy* **1997**, *3*, 469–475.
- (39) Hanssen, L. M.; Zhu, C. In *Handbook of Vibrational Spectroscopy*; Chalmers, J. M., Griffiths, P. R., Eds.; John Wiley & Sons Ltd: Chichester, 2002; pp 881–890.
- (40) Rothman, L. S.; Gordon, I. E.; Babikov, Y.; Barbe, A.; Benner, D. C.; Bernath, P. F.; Birk, M.; Bizzocchi, L.; Boudon, V.; Brown, L. R.; Campargue, A.; Chance, K.; Cohen, E. A.; Coudert, L. H.; Devi, V. M.; Drouin, B. J.; Fayt, A.; Flaud, J. M.; Gamache, R. R.; Harrison, J. J.; Hartmann, J. M.; Hill, C.; Hodges, J. T.; Jacquemart, D.; Jolly, A.; Lamouroux, J.; Le Roy, R. J.; Li, G.; Long, D. A.; Lyulin, O. M.; Mackie, C. J.; Massie, S. T.; Mikhailenko, S.; Muller, H. S. P.; Naumenko, O. V.; Nikitin, A. V.; Orphal, J.; Perevalov, V.; Perrin, A.; Polovtseva, E. R.; Richard, C.; Smith, M. A. H.; Starikova, E.; Sung, K.; Tashkun, S.; Tennyson, J.; Toon, G. C.; Tyuterev, V. G.; Wagner, G. J. *Quant. Spectrosc. Radiat. Transfer* **2013**, *130*, 4–50.
- (41) Venyaminov, S. Y.; Prendergast, F. G. *Anal. Biochem.* **1997**, *248*, 234–245.
- (42) Praprotnik, M.; Janežič, D.; Mavri, J. J. *Phys. Chem. A* **2004**, *108*, 11056–11062.
- (43) Freda, M.; Piluso, A.; Santucci, A.; Sassi, P. *Appl. Spectrosc.* **2005**, *59*, 1155–1159.
- (44) Amunson, K. E.; Anderson, B. A.; Kubelka, J. *Appl. Spectrosc.* **2011**, *65*, 1307–1313.
- (45) Smith, R.; Ellis, B. *Polymers*, 2nd ed.; CRC Press: Boca Raton, 2008; pp 333–1027.